

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program


APR 13 2010

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130298
Name		TIM TUCKWOOD	ID# 3446 Date 4-7-10
A	Agency	Alaska State Troopers	Phone # 895-4800
Instrument Location		Arctic Man Event	
Alco S/N		1B	Target Value 080 High Pressure 950
B	Alco Test Values		072 070
		1 st Alco	2 nd Alco
Signature		TIM TUCKWOOD	
(OVER)			


(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.


Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 26th day of April, 2010.


Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office

(Notary Seal Stamp)



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130298 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130298

APRIL 07, 2010 ✓

OPERATOR'S NAME:

TUCKWOOD TIM

OPERATOR'S NUMBER: 3446

SUBJECT'S LAST NAME:

VERIF CAL

SUBJECT'S FIRST NAME/MI :

VERIF OF CAL

O.L. #: 12345

DEPT/AGENCY: DJT0

CASE/REPORT: 09876

TEST TYPE: U

ALCO TARGET VALUE: .000 ✓

ALCO S/N: 10

— BREATH ANALYSIS —

.000 ADJUSTED FOR 26.45 in ✓
ALCO TARGET .070 19:11
BLANK TEST .000 19:12
INTERNAL STANDARD VERIFIED 19:12
ALCO TV 26.45 in .072 19:13
BLANK TEST .000 19:13
SUBJECT SAMPLE .000 19:14
BLANK TEST .000 19:14
ALCO TV 26.45 in .070 19:15
BLANK TEST .000 19:16

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130298 ✓

APRIL 07, 2010 ✓
TIME 19:09

— DIAGNOSTIC CHECK —

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 01/08/09
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 41c
BAROMETER: 26.45 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!""&%'()*+,-./0123456789:;<=>?@ABCDEFGHIJ
KLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmnopqrstuvwxyz{|}~